



**University of
Zurich**^{UZH}

**Zurich Open Repository and
Archive**

University of Zurich
University Library
Strickhofstrasse 39
CH-8057 Zurich
www.zora.uzh.ch

Year: 2018

Fiber-Reinforced Composites for Dental Applications

Scribante, Andrea ; Vallittu, Pekka K ; Özcan, Mutlu

DOI: <https://doi.org/10.1155/2018/4734986>

Posted at the Zurich Open Repository and Archive, University of Zurich

ZORA URL: <https://doi.org/10.5167/uzh-162776>

Journal Article

Accepted Version

Originally published at:

Scribante, Andrea; Vallittu, Pekka K; Özcan, Mutlu (2018). Fiber-Reinforced Composites for Dental Applications. BioMed Research International, 2018:4734986.

DOI: <https://doi.org/10.1155/2018/4734986>

Editorial

Special issue:

Fiber-Reinforced Composites for Dental Applications

Authors:

Andrea Scribante

Unit of Orthodontics and Pediatric Dentistry, Section of Dentistry, Department of Clinical, Surgical, Diagnostic and Pediatric Sciences, University of Pavia, Italy
andrea.scribante@unipv.it

Pekka K. Vallittu

Department of Biomaterial Science and Turku Clinical Biomaterials Centre-TCBC, Institute of Dentistry, University of Turku, Turku, Finland
City of Turku, Welfare Division, Turku, Finland
pekka.vallittu@utu.fi

Mutlu Özcan

University of Zurich, Center for Dental and Oral Medicine, Dental Materials Unit, Clinic for Fixed and Removable Prosthodontics and Dental Materials Science, Zurich, Switzerland
mutluozcan@hotmail.com

Fiber-reinforced composites (FRCs) are composite materials with three different components: the matrix (continuous phase), the fibers (dispersed phase) and the zone in between (interphase). FRC materials present high stiffness and strength per weight when compared with other structural materials along with adequate toughness. FRCs have been used for numerous applications in various engineering and biomedical fields for a long time. The reinforcement of dental resins with either short or long fibers on the other hand, has been described in literature for more than 40 years [Vallittu and Lassila, 1992]. FRCs based on carbon, polyaramid, polyethylene, and glass have been largely studied and among all, glass fibers of various compositions are more commonly applied as restorative and prosthetic materials [Vallittu and Sevelius, 2000; Saker and Özcan 2015].

FRCs have been intensively investigated with a particular emphasis on mechanical properties such as fracture toughness, compressive strength, load-bearing capacity [Garoushi et al., 2011], flexural strength [Cacciafesta et al., 2007; Cacciafesta et al., 2008], fatigue resistance [Foek et al., 2013], fracture strength [Kumbuloglu et al., 2008] or on the effect of layer thickness [Bjelic-Donova et al., 2015], bacterial adhesion [Akalin-Evren et al., 2014], adhesion of fibers for various dental applications [Scribante et al., 2006; Bell et al., 2004; Sfondrini et al., 2011]. From clinical perspective, FRCs have been investigated for different clinical applications in prosthodontics, such as replacement of missing teeth by resin-bonded adhesive fixed dental prostheses of various kinds [Karaman et al., 2002; Kumbuloglu and Özcan, 2015], reinforcement elements of dentures or pontics [Garoushi et al., 2012] and direct construction of posts and cores [Zorba and Ozcan, 2007]. In other disciplines of dentistry, such as orthodontics FRCs have been suggested as active and passive orthodontic applications, (i.e. anchorage or en-masse movement units) and post-orthodontic tooth retention [Scribante et al., 2011; Sfondrini et al., 2014] and in periodontology for splinting mobile teeth in an attempt to prolong tooth extraction [Kumbuloglu et al., 2011].

With the introduction of new technologies, nanofillers, resin matrices, fibers, adhesion protocols, application techniques the design principles of FRC devices need further understanding which open new fields of research both preclinically and clinically [Sfondrini et al., 2014; Scribante et al., 2015]. On the basis of these considerations, BioMed Research International prepared the present Special Issue on “Fiber Reinforced Composites for Dental Applications” in an attempt to explore these new variables related to FRCs.

Guest Editors do hope that this Special Issue would be interesting for the readers of the journal and wish that the present work could help both clinicians and researchers to understand FRC applications and properties. Finally, Guest Editors would like to thank the Editorial Board of BioMed Research International for the invitation to prepare this Special Issue. A special thank is also addressed to Sam Rose for kind suggestions during call-for-papers and to Sara Ashraf for excellent assistance during manuscript management track.

References

1. Vallittu PK, Lassila VP. Reinforcement of acrylic resin denture base material with metal or fibre strengtheners. *J Oral Rehabil* 1992;19:225-230
2. Vallittu PK, Sevelius C. Resin-bonded, glass fiber reinforced composite fixed partial dentures - A clinical study. *J Prosthet Dent* 2000;84:413-418
3. Saker S, Özcan M. Retentive strength of fiber-reinforced composite posts with composite resin cores: Effect of remaining coronal structure and root canal dentin conditioning protocols. *J Prosthet Dent*. 2015;114:856-861
4. Garoushi S, Vallittu PK, Lassila LVJ. Fracture toughness, compressive strength and load-bearing capacity of short glass fiber-reinforced composite resin. *Chin J Dent Res* 2011;14:15-9.
5. Cacciafesta V, Sfondrini MF, Lena A, Scribante A, Vallittu PK, Lassila LV. Flexural strengths of fiber-reinforced composites polymerized with conventional light-curing and additional postcuring. *Am J Orthod Dentofacial Orthop* 2007;132:524-527
6. Cacciafesta V, Sfondrini MF, Lena A, Scribante A, Vallittu PK, Lassila LV. Force levels of fiber-reinforced composites and orthodontic stainless steel wires: a 3-point bending test. *Am J Orthod Dentofacial Orthop* 2008;133:410-413
7. Foek DL, Yetkiner E, Özcan M. Fatigue resistance, debonding force, and failure type of fiber-reinforced composite, polyethylene ribbon-reinforced, and braided stainless steel wire lingual retainers in vitro. *Korean J Orthod* 2013;43:186-192
8. Kumbuloglu O, Özcan M, User A. Fracture strength of direct surface-retained fixed partial dentures: effect of fiber reinforcement versus the use of particulate filler composites only. *Dent Mater J* 2008;27:195-202
9. Bijelic-Donova J, Garoushi S, Vallittu PK, Lassila LVJ. Oxygen inhibition layer of composite resins: the effect of layer thickness and surface layer treatment on the interlayer bond strength. *Eur J Oral Sci* 2015;123:53-60

10. Akalın-Evren B, Kulak-Özkan Y, Özcan M, Kadir T. Candida albicans adhesion on reinforced polymethylmethacrylate denture resin: effect of fibre architecture and exposure to saliva. *Gerodontology* 2014;31:194-201.
11. Scribante A, Cacciafesta V, Sfondrini MF. Effect of various adhesive systems on the shear bond strength of fiber-reinforced composite. *Am J Orthod Dentofacial Orthop* 2006;130:224-227.
12. Sfondrini MF, Cacciafesta V, Scribante A. Shear bond strength of fibre-reinforced composite nets using two different adhesive systems. *Eur J Orthod* 2011;33:66-70
13. LeBell A-M, Tanner J, Lassila LVJ, Kangasniemi I, Vallittu PK. Bonding of composite resin luting cement to fibre-reinforced composite root canal post. *J Adhesive Dent* 2004;6:319-325
14. Karaman AI, Kir N, Belli S. Four applications of reinforced polyethylene fiber material in orthodontic practice. *Am J Orthod Dentofacial Orthop* 2002;121:650-654
15. Kumbuloglu O, Özcan M. Clinical survival of indirect, anterior 3-unit surface retained fiber-reinforced composite fixed dental prosthesis: 7.5-year follow-up. *J Dent* 2015;43:656-663
16. Garoushi S, Lassila L, Vallittu PK. Single visit replacement of maxillary canine using fiber-reinforced composite resin. *J Contemp Dent Pract* 2012;13:125129
17. Zorba YO, Ozcan E. Reattachment of coronal fragment using fiber-reinforced post: a case report. *Eur J Dent* 2007;1:174-178
18. Scribante A, Sfondrini MF, Broggin S, D'Allocco M, Gandini P. efficacy of esthetic retainers: clinical comparison between multistranded wires and direct-bond glass fiber-reinforced composite splints. *Int J Dent* 2011;2011:548356
19. Sfondrini MF, Fraticelli D, Castellazzi L, Scribante A, Gandini P. Clinical evaluation of bond failures and survival between mandibular canine-to-canine retainers made of flexible spiral wire and fiber-reinforced composite. *J Clin Exp Dent* 2014;6:e145-149
20. Kumbuloglu O, Saracoglu A, Özcan M. Pilot study of unidirectional e-glass fiber-reinforced composite resin splints: up to 4.5-year clinical follow-up. *J. Dent.* 2011;39:871-877
21. Scribante A, Massironi S, Pieraccini G, Vallittu P, Lassila L, Sfondrini MF, Gandini P. Effects of nanofillers on mechanical properties of fiber-reinforced composites polymerized with light-curing and additional postcuring. *J Appl Biomater Funct Mater* 2015;13:e296-299

22. Sfondrini MF, Massironi S, Pieraccini G, Scribante A, Vallittu PK, Lassila LV, Gandini P. Flexural strengths of conventional and nanofilled fiber-reinforced composites: a three-point bending test. *Dent Traumatol* 2014;30:32-35